

## QA/QC Results of Laboratory Analysis of Total Suspended Solids

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
11/5/2013	99.4	FC1S-1	3.17	FS1S-2	100.0
	102.8	FS1S-3	0.00	FE7M-1	97.9
	99.2	FE7M-2	2.67	FG3M-3	103.8
	102.4	FG3B-1	2.60	FS2B-2	101.9
	98.3	FS2B-3	2.82	FS3B-3	108.3
	105.7	EC1S-1	0.00	ES1S-2	94.2
	93.3	ES1S-3	3.17	EE7M-1	98.1
	105.8	EE7M-2	2.90	EG3M-3	96.0
	97.0	EG3B-1	0.00	ES2B-2	103.8
	106.6	ES2B-3	0.00	ES3B-3	91.8

Note: (\*) % Recovery of QC sample should be between 80% to 120%.  
 (#) % Error of Sample Duplicate should be between 0% to 10%.  
 (@) % Recovery of Sample Spike should be between 80% to 120%.  
 (\*\*) % Error of Sample Duplicate >10% but invalid due to sample results less than ML

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
11/7/2013	94.8	FC1S-1	3.28	FG1S-2	100.0
	102.3	FG1M-1	2.99	FG3M-2	100.0
	106.5	FG3B1	2.74	FG2B-2	104.1
	95.9	FS3S-1	0.00	ES1S-2	98.1
	92.6	ES1M-1	0.00	EE1M-2	94.1
	106.5	EF1B1	2.67	ES2B-2	101.9
		100.2	EG2S1	0.00	ES3B-2

Note: (\*) % Recovery of QC sample should be between 80% to 120%.  
 (#) % Error of Sample Duplicate should be between 0% to 10%.  
 (@) % Recovery of Sample Spike should be between 80% to 120%.  
 (\*\*) % Error of Sample Duplicate >10% but invalid due to sample results less than ML

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
11/9/2013	106.8	FC1S-1	2.90	FG1S-2	105.9
	106.0	FG1M-1	2.99	FG3M-2	100.0
	94.9	FG3B1	0.00	FG2B-2	106.0
	92.3	FS3S-1	2.99	ES1S-2	103.8
	105.7	ES1M-1	0.00	EE1M-2	93.9
	94.3	EF1B1	2.82	ES2B-2	103.8
		93.0	EG2S1	3.17	ES3B-2

Note: (\*) % Recovery of QC sample should be between 80% to 120%.  
 (#) % Error of Sample Duplicate should be between 0% to 10%.  
 (@) % Recovery of Sample Spike should be between 80% to 120%.  
 (\*\*) % Error of Sample Duplicate >10% but invalid due to sample results less than ML