



Internal Calibration Report of Dissolved Oxygen Meter

| | |
|---|--|
| Equipment Ref. No. : <u>ET/EW/008/005</u> | Manufacturer : <u>YSI</u> |
| Model No. : <u>Pro 2030</u> | Serial No. : <u>12A 100353</u> |
| Date of Calibration : <u>25/08/2012</u> | Calibration Due Date : <u>24/11/2012</u> |

Temperature Verification

Ref. No. of Reference Thermometer : ET/0521/001
 Ref. No. of Water Bath : ---

| | | Temperature (°C) | | |
|-------------------------------|----------|------------------|------------|------|
| Reference Thermometer reading | Measured | 20.2 | Corrected | 19.8 |
| DO Meter reading | Measured | 19.7 | Difference | 0.1 |

Standardization of sodium thiosulphate (Na₂S₂O₃) solution

| | | | |
|---|-------------------|---|--------------------|
| Reagent No. of Na ₂ S ₂ O ₃ titrant | CPE/012/4.5/001/5 | Reagent No. of 0.025N K ₂ Cr ₂ O ₇ | CPE/012/4.4/001/12 |
| | | Trial 1 | Trial 2 |
| Initial Vol. of Na ₂ S ₂ O ₃ (ml) | | 0.00 | 0.00 |
| Final Vol. of Na ₂ S ₂ O ₃ (ml) | | 40.10 | 40.05 |
| Vol. of Na ₂ S ₂ O ₃ used (ml) | | 40.10 | 40.05 |
| Normality of Na ₂ S ₂ O ₃ solution (N) | | 0.02494 | 0.02497 |
| Average Normality (N) of Na ₂ S ₂ O ₃ solution (N) | | 0.02496 | |
| Acceptance criteria, Deviation | | Less than ± 0.001N | |

Calculation: Normality of Na₂S₂O₃, N = 1 / ml Na₂S₂O₃ used

Linearity Checking

Determination of dissolved oxygen content by Winkler Titration *

| Purging Time (min) | 2 | | 5 | | 10 | |
|---|---------------------|-------|---------------------|------|---------------------|-------|
| | 1 | 2 | 1 | 2 | 1 | 2 |
| Initial Vol. of Na ₂ S ₂ O ₃ (ml) | 0.00 | 11.20 | 22.20 | 0.00 | 7.60 | 12.30 |
| Final Vol. of Na ₂ S ₂ O ₃ (ml) | 11.20 | 22.20 | 29.90 | 7.60 | 12.30 | 17.20 |
| Vol. (V) of Na ₂ S ₂ O ₃ used (ml) | 11.20 | 11.00 | 7.70 | 7.60 | 4.70 | 4.90 |
| Dissolved Oxygen (DO), mg/L | 7.50 | 7.37 | 5.16 | 5.09 | 3.15 | 3.28 |
| Acceptance criteria, Deviation | Less than + 0.3mg/L | | Less than + 0.3mg/L | | Less than + 0.3mg/L | |

Calculation: DO (mg/L) = V x N x 8000/298

| Purging time, min | DO meter reading, mg/L | | | Winkler Titration result *, mg/L | | | Difference (%) of DO Content |
|-------------------------------|------------------------|------|---------|----------------------------------|------|---------|------------------------------|
| | 1 | 2 | Average | 1 | 2 | Average | |
| 2 | 7.51 | 7.60 | 7.56 | 7.50 | 7.37 | 7.44 | 1.60 |
| 5 | 5.21 | 5.20 | 5.21 | 5.16 | 5.09 | 5.13 | 1.55 |
| 10 | 3.19 | 3.25 | 3.22 | 3.15 | 3.28 | 3.22 | 0.00 |
| Linear regression coefficient | | | | 0.99990 | | | |



Internal Calibration Report of Dissolved Oxygen Meter

Zero Point Checking

| | |
|------------------------|------|
| DO meter reading, mg/L | 0.00 |
|------------------------|------|

Salinity Checking

| | | | |
|-----------------------------|--------------------|-----------------------------|--------------------|
| Reagent No. of NaCl (10ppt) | CPE/012/4.7/001/28 | Reagent No. of NaCl (30ppt) | CPE/012/4.8/001/28 |
|-----------------------------|--------------------|-----------------------------|--------------------|

*Determination of dissolved oxygen content by Winkler Titration ***

| Salinity (ppt) | 10 | | 30 | |
|---|---------------------|-------|---------------------|-------|
| | 1 | 2 | 1 | 2 |
| Initial Vol. of Na ₂ S ₂ O ₃ (ml) | 0.00 | 11.50 | 23.20 | 33.90 |
| Final Vol. of Na ₂ S ₂ O ₃ (ml) | 11.50 | 23.20 | 33.90 | 44.40 |
| Vol. (V) of Na ₂ S ₂ O ₃ used (ml) | 11.50 | 11.70 | 10.70 | 10.50 |
| Dissolved Oxygen (DO), mg/L | 7.71 | 7.84 | 7.17 | 7.04 |
| Acceptance criteria, Deviation | Less than + 0.3mg/L | | Less than + 0.3mg/L | |

Calculation: $DO (mg/L) = V \times N \times 8000/298$

| Salinity (ppt) | DO meter reading, mg/L | | | Winkler Titration result**, mg/L | | | Difference (%) of DO Content |
|----------------|------------------------|------|---------|----------------------------------|------|---------|------------------------------|
| | 1 | 2 | Average | 1 | 2 | Average | |
| 10 | 7.7 | 7.65 | 7.68 | 7.71 | 7.84 | 7.78 | 1.29 |
| 30 | 7.13 | 7.05 | 7.09 | 7.17 | 7.04 | 7.11 | 0.28 |

Acceptance Criteria

- (1) Difference between temperature readings from temperature sensor of DO probe and reference thermometer : < 0.5 °C
- (2) Linear regression coefficient : >0.99
- (3) Zero checking: 0.0mg/L
- (4) Difference (%) of DO content from the meter reading and by winkler titration : within ± 5%

The equipment complies # / does not comply # with the specified requirements and is deemed acceptable # / unacceptable # for use.

Delete as appropriate

Calibrated by

:

Approved by :



Performance Check of Salinity Meter

Equipment Ref. No. : ET/EW/008/005 Manufacturer : YSI
Model No. : Pro 2030 Serial No. : 12A 100353
Date of Calibration : 25/08/2012 Due Date : 24/11/2012

Ref. No. of Salinity Standard used (30ppt)


S/001/3


| Salinity Standard (ppt) | Measured Salinity (ppt) | Difference % |
|-------------------------|-------------------------|--------------|
| 30.0 | 30.2 | 0.66 |

Acceptance Criteria

Difference : <10 %

The salinity meter complies * / ~~does not comply~~ * with the specified requirements and is deemed acceptable * / ~~unacceptable~~ * for use. Measurements are traceable to national standards.

Checked by : 

Approved by : 



Performance Check of Turbidimeter


Equipment Ref. No. : ET/0505/008 Manufacturer : HACH
Model No. : 2100Q Serial No. : 10030 C 001191
Date of Calibration : 02/08/2012 Due Date : 01/11/2012

| Gelex Vial Std | Theoretical Value (NTU) | Measured Value (NTU) | Difference % |
|----------------|-------------------------|----------------------|--------------|
| 0-10 NTU | 5.70 | 5.62 | 1.41 |
| 10-100 NTU | 52.1 | 52.7 | 1.15 |
| 100-1000 NTU | 547 | 539 | 1.47 |

Acceptance Criteria

Difference : <5 %

The salinity meter complies * / ~~does not comply~~ * with the specified requirements and is deemed acceptable * / ~~unacceptable~~ * for use. Measurements are traceable to national standards.

Checked by :  Approved by : 