

QA/QC Results of Laboratory Analysis of Total Suspended Solids

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
9/24/2012	104.7	FE4S-1	11.76**	FG6 S-2	93.9
	103.0	FG6 M-1	8.70	FG5 B-2	100.0
	93.9	EE4S-1	0.00	EG6 S-2	94.3
	105.7	EG6 M-1	8.00	EG5 B-2	98.1

Note: (*) % Recovery of QC sample should be between 80% to 120%.
 (#) % Error of Sample Duplicate should be between 0% to 10%.
 (@) % Recovery of Sample Spike should be between 80% to 120%.
 (**) % Error of Sample Duplicate >10% but invalid due to sample results less than ML

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
9/26/2012	106.5	FE4S-1	0.00	FG6 S-2	103.9
	92.1	FG6 M-1	8.70	FG5 B-2	106.1
	99.6	EE4S-1	8.70	EG6 S-2	93.9
	100.2	EG6 M-1	9.52	EG5 B-2	96.0

Note: (*) % Recovery of QC sample should be between 80% to 120%.
 (#) % Error of Sample Duplicate should be between 0% to 10%.
 (@) % Recovery of Sample Spike should be between 80% to 120%.
 (**) % Error of Sample Duplicate >10% but invalid due to sample results less than ML

Sampling Date	QC Sample	Sample Duplicate		Sample Spike	
	% Recovery *	Sample ID	% Error #	Sample ID	% Recovery @
9/28/2012	103.6	FE4S-1	0.00	FG6 S-2	93.8
	94.4	FG6 M-1	0.00	FG5 B-2	91.7
	107.6	EE4S-1	0.00	EG6 S-2	98.1
	100.8	EG6 M-1	0.00	EG5 B-2	98.0

Note: (*) % Recovery of QC sample should be between 80% to 120%.
 (#) % Error of Sample Duplicate should be between 0% to 10%.
 (@) % Recovery of Sample Spike should be between 80% to 120%.
 (**) % Error of Sample Duplicate >10% but invalid due to sample results less than ML